



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Golden Falls Trading 429 (Pty) Ltd  
T/A Coral-i Services**

**508 Nupen Crescent, Halfway House  
Midrand Gauteng, South Africa 1685**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 30 March 2026

Certificate Number: L2448



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Golden Falls Trading 429 (Pty) Ltd**  
**T/A Coral-i Services**  
 508 Nupen Crescent  
 Halfway House  
 Midrand Gauteng, South Africa 1685  
 Werner Muller +27 (11) 315 5500

**CALIBRATION**

Valid to: **March 30, 2026**

Certificate Number: **L2448**

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Generate <sup>1</sup>	(0 to 320) $\mu$ A (0.32 to 3.2) mA (3.2 to 32) mA (32 to 320) mA (0.32 to 3.2) A (3.2 to 10.5) A (10.5 to 20) A	30 nA + 0.014 % of reading 0.1 pA + 0.014 % of reading 1.2 $\mu$ A + 0.014 % of reading 13 $\mu$ A + 0.016 % of reading 160 $\mu$ A + 0.06 % of reading 1.5 mA + 0.055 % of reading 5.4 mA + 0.055 % of reading	Comparison to Wavetek 9100 Universal Calibration System
DC Current - Measure <sup>1</sup>	(0 to 10) mA (10 to 100) mA (0.1 to 1) A (1 to 3) A	23 $\mu$ A + 0.05 % of reading 8 $\mu$ A + 0.05 % of reading 0.15 mA + 0.1 % of reading 0.8 mA + 0.12 % of reading	Comparison to HP 34401A Multimeter
AC Current (Generate) <sup>1</sup>	(32 to 320) $\mu$ A 10 Hz to 3 kHz (3 to 10) kHz (10 to 20) kHz (0.32 to 3.2) mA 10 Hz to 3 kHz (3 to 10) kHz (10 to 20) kHz (3.2 to 32) mA 10 Hz to 3 kHz (3 to 10) kHz (10 to 20) kHz	4 $\mu$ A + 0.07 % of reading 4.1 $\mu$ A + 0.1 % of reading 4.6 $\mu$ A + 0.2 % of reading 4 $\mu$ A + 0.07 % of reading 4.1 $\mu$ A + 0.1 % of reading 4.6 $\mu$ A + 0.2 % of reading 8.8 $\mu$ A + 0.07 % of reading 11 $\mu$ A + 0.1 % of reading 17 $\mu$ A + 0.2 % of reading	Comparison to Wavetek 9100 Universal Calibration System
AC Current (Generate) <sup>1</sup>	(32 to 320) mA 10 Hz to 3 kHz (3 to 10) kHz (10 to 20) kHz	88 $\mu$ A + 0.08 % of reading 97 $\mu$ A + 0.1 % of reading 0.11 mA + 0.2 % of reading	Comparison to Wavetek 9100 Universal Calibration System

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current (Generate) <sup>1</sup>	(0.32 to 3.2) A 10 Hz to 3 kHz (3 to 10) kHz (3.2 to 10.5) A 10 Hz to 3 kHz (3 to 10) kHz	0.82 mA + 0.1 % of reading 3 mA + 0.25 % of reading  7 mA + 0.2 % of reading 13 mA + 0.5 % of reading	Comparison to Wavetek 9100 Universal Calibration System
AC Current (Measure) <sup>1</sup>	(0 to 1) A 10 Hz to 5 kHz (1 to 3) A 10 Hz to 5 kHz	0.61 mA + 0.1 % of reading  2.4 mA + 0.15 % of reading	Comparison to HP 34401A Multimeter
Resistance (Generate) <sup>1</sup>	(0 to 40) Ω (40 to 400) Ω (400 to 4) kΩ (4 to 40) kΩ (40 to 400) kΩ (0.4 to 4) MΩ (4 to 40) MΩ	58 mΩ + 0.1 % of reading 0.12 μΩ + 0.035 % of reading 0.25 μΩ + 0.035 % of reading 2.4 Ω + 0.025 % of reading 24 Ω + 0.025 % of reading 0.23 kΩ + 0.04 % of reading 2.3 kΩ + 0.05 % of reading	Comparison to Wavetek 9100 Universal Calibration System
Resistance (Measure) <sup>1</sup>	(0 to 100) Ω (0.1 to 1) kΩ (1 to 10) kΩ (10 to 100) kΩ (0.1 to 1) MΩ (1 to 10) MΩ (10 to 100) MΩ	0.05 Ω + 0.01 % of reading 0.02 Ω + 0.01 % of reading 0.15 Ω + 0.01 % of reading 1.7 Ω + 0.01 % of reading 23 Ω + 0.01 % of reading 0.42 kΩ + 0.04 % of reading 2.3 kΩ + 0.8 % of reading	Comparison to HP 34401A Multimeter
DC Voltage (Generate) <sup>1</sup>	(0 to 320) mV (0.32 to 3.2) V (3.2 to 32) V (32 to 320) V (320 to 1 050) V	5.5 μV + 0.006 % of reading 55 μV + 0.006 % of reading 550 μV + 0.006 5 % of reading 5.7 mV + 0.006 5 % of reading 30 mV + 0.006 % of reading	Comparison to Wavetek 9100 Universal Calibration System
DC Voltage (Measure) <sup>1</sup>	(0 to 100) mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1 000) V	0.004 mV + 0.005 % of reading 0.01 mV + 0.004 % of reading 0.07 mV + 0.003 5 % of reading 0.93 mV + 0.004 5 % of reading 14 mV + 0.004 5 % of reading	Comparison to HP 34401A Multimeter
AC Voltage (Generate) <sup>1</sup>	(0 to 10) mV 10 Hz to 3 kHz (3 to 10) kHz (10 to 32) mV 10 Hz to 3 kHz (3 to 10) kHz	490 μV + 0.08 % of reading 630 μV + 0.08 % of reading  230 μV + 0.08 % of reading 250 μV + 0.08 % of reading	Comparison to Wavetek 9100 Universal Calibration System, HP 34401A Multimeter

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
AC Voltage (Generate) <sup>1</sup>	(32 to 320) mV 10 Hz to 3 kHz (3 to 10) kHz	100 μV + 0.08 % of reading 110 μV + 0.08 % of reading	Comparison to Wavetek 9100 Universal Calibration System, HP 34401A Multimeter		
	(0.32 to 3.2) V 10 Hz to 3 kHz (3 to 10) kHz	250 μV + 0.08 % of reading 320 μV + 0.08 % of reading			
	(3.2 to 32) V 10 Hz to 3 kHz (3 to 10) kHz	2.3 mV + 0.08 % of reading 3 mV + 0.1 % of reading			
	(32 to 105) V 10 Hz to 3 kHz (3 to 10) kHz	8 mV + 0.08 % of reading 10 mV + 0.1 % of reading			
	(105 to 320) V 40 Hz to 1 kHz (1 to 3) kHz (2 to 10) kHz	23 mV + 0.09 % of reading 23 mV + 0.12 % of reading 38 mV + 0.12 % of reading			
	(320 to 800) V 40 Hz to 1 kHz (1 to 3) kHz (2 to 10) kHz	90 mV + 0.09 % of reading 91 mV + 0.12 % of reading 140 mV + 0.12 % of reading			
	AC Voltage (Measure) <sup>1</sup>	(0 to 100) mV 10 Hz to 20 kHz		0.05 mV + 0.06 % of reading	Comparison to Wavetek 9100 Universal Calibration System, HP 34401A Multimeter
		(0.1 to 1) V 10 Hz to 20 kHz		0.35 mV + 0.06 % of reading	
		(1 to 10) V 10 Hz to 20 kHz		3.5 mV + 0.06 % of reading	
		(10 to 100) V 10 Hz to 20 kHz		35 mV + 0.06 % of reading	
		(10 to 100) V 10 Hz to 20 kHz		270 mV + 0.06 % of reading	
		Oscilloscopes <sup>1</sup> Vertical System – 50 Ω Vertical System – 1 MΩ		1 mV to 2 V/div 1 mV to 20 V/div	
Oscilloscopes Horizontal System – 50 Ω	2 ns to 5 s/div		0.31 % of reading		
Oscilloscopes <sup>1</sup> Bandwidth – 50 Ω 1 mV to 2 V/div	50 kHz to 100 MHz	5.4 % of bandwidth	Comparison to Wavetek 9100 Universal Calibration System		
	50 kHz to 250 MHz	6.2 % of bandwidth			
	50 kHz to 600 MHz	7.7 % of bandwidth			

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Attenuation <sup>1</sup>	(0 to 11) dB, 1 dB steps 10 MHz to 10 GHz (0 to 70) dB, 10 dB steps (10 to 18) GHz	0.35 dB + 0.005 dB/dB  0.65 dB + 0.015 dB/dB	Comparison to Step attenuator, power meter, power senso
Voltage Reflection Coefficient – (0 to 1) (N-M)  VRC (N-F)	10 MHz to 2 GHz (2 to 10) GHz (10 to 18) GHz  10 MHz to 2 GHz (2 to 10) GHz (10 to 18) GHz	0.016 + 0.013ρ <sup>2</sup> + 0.067ρ 0.033 + 0.043ρ <sup>2</sup> + 0.067ρ 0.085 + 0.062ρ <sup>2</sup> + 0.067ρ  0.016 + 0.013ρ <sup>2</sup> + 0.067ρ 0.029 + 0.046ρ <sup>2</sup> + 0.067ρ 0.036 + 0.076ρ <sup>2</sup> + 0.067ρ	Comparison to Scalar network analyzer
RF Power – Measure <sup>1</sup>	1 mW  50 MHz	  1 % of reading	Comparison to Step attenuator, power meter, power senso
RF Power – Measure <sup>1</sup>	(-25 to 13) dBm 100 kHz to 10 MHz (-50 to 13) dBm 10 MHz to 50 MHz 50 MHz to 6 GHz 6 GHz to 18 GHz 18 GHz to 26 GHz 26 GHz to 36 GHz 36 GHz to 40 GHz	0.33 dB  0.15 dB 0.14 dB 0.21 dB 0.27 dB 0.34 dB 0.53 dB	Comparison to RF Power meter w/Power Sensors. UUT SWR ≤1.5
RF Power – Measure <sup>1</sup>	(13 to 50) dBm 10 MHz to 1 GHz	0.21 dB	Comparison to RF Power meter w/Power Sensors, Attenuator
RF Power – Source <sup>1</sup>	1 mW  50 MHz	  1 % of reading	Comparison to Output of RF Power Meter
RF Power – Source <sup>1</sup>	(-25 to 13) dBm 100 kHz to 10 MHz (-50 to 13) dBm 10 MHz to 50 MHz 50 MHz to 6 GHz (6 GHz to 18) GHz (18 GHz to 26) GHz (26 GHz to 36) GHz (36 GHz to 40) GHz (13 to 50) dBm 200 MHz to 1 GHz	0.46 dB  0.17 dB  0.16 dB 0.26 dB 0.32 dB 0.39 dB 0.66 dB 0.31 dB	Comparison to RF Power meter w/Power Sensors, RF Signal Generators, Power Amplifier

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Sensors – Calibration Factor <sup>1</sup>	100 kHz to 300 kHz 300 kHz to 50 MHz 50 MHz to 6 GHz 6 GHz to 18 GHz 18 GHz to 26 GHz 26 GHz to 36 GHz 36 GHz to 40 GHz	5.3 % of reading 1.7 % of reading 2 % of reading 3.4 % of reading 5 % of reading <sup>2</sup> 6.7 % of reading <sup>2</sup> 10 % of reading <sup>2</sup>	Comparison to Reference power sensor and power meter with RF Generator
Amplitude Modulation – Measure & Generate <sup>1</sup> CW: 10 MHz to 1.3 GHz CW: 150 kHz to 10 MHz	Rate: 50 Hz to 50 kHz Depth (5 to 99) %Depth  Rate: 50 Hz to 10 kHz Depth (5 to 99) %Depth	1.7 % of reading  2.6 % of reading	Comparison to Modulation Meter w/ RF Signal Generator
Frequency Modulation – Measure & Generate <sup>1</sup> CW: 10 MHz to 1.3 GHz CW: 250 kHz to 10 MHz	Rate: 50 Hz to 100 kHz 400 kHz Peak Deviation  Rate: 20 Hz to 10 kHz 40 kHz Peak Deviation	1.3 % of reading  2.4 % of reading	Comparison to Modulation Meter w/ RF Signal Generator
Distortion <sup>1</sup>	(-85 to 0) dB 20 Hz to 20 kHz 20 kHz to 100 kHz	1.4 dB 2.5 dB	Comparison to Audio Analyzer

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers – Outside, Inside Measurements	(0 to 150) mm	17 µm	Comparison to Multifunction Gauge Block
Micrometers	(0 to 150) mm	12 µm	Comparison to Multifunction Gauge Block
Dial Gages	(0 to 10) mm	4.5 µm	Comparison to Multifunction Gauge Block

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Torque Tools	(2 to 40) N·m	1.8 % of reading	Comparison to Torque transducer and readout

**Photometry and Radiometry**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Time Domain Reflectometer – (OTDR): Distance – Wavelengths: (1310, 1550, 1625) nm Refractive Index 1.46	Fiber Length: 18.8 km	0.5 m + 1 sampling interval	Comparison to Single mode Optical Delay Line
Optical Power Level (850 to 1 650) nm	(-50 to +20) dBm	0.3 dB + 46 nW	Comparison to Optical Power meter & head


**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Measure <sup>3</sup>	10 Hz to 100 kHz 100 kHz to 1 MHz (1 to 10) MHz 10 MHz to 40 GHz	$5.8 \times 10^{-8} \cdot f + 0.0008 \text{ Hz}$ $5.8 \times 10^{-8} \cdot f + 0.0082 \text{ Hz}$ $5.8 \times 10^{-8} \cdot f + 0.082 \text{ Hz}$ $5.8 \times 10^{-8} \cdot f + 0.82 \text{ Hz}$	Comparison to GPS Receiver, Frequency Counter
Frequency – Generate <sup>3</sup>	10 Hz to 100 kHz 100 kHz to 1 MHz (1 to 10) MHz 10 MHz to 40 GHz	$5.8 \times 10^{-8} \cdot f + 0.0010 \text{ Hz}$ $5.8 \times 10^{-8} \cdot f + 0.01 \text{ Hz}$ $5.8 \times 10^{-8} \cdot f + 0.1 \text{ Hz}$ $5.8 \times 10^{-8} \cdot f + 1.0 \text{ Hz}$	Comparison to GPS Receiver, Frequency Counter, Signal Generator

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. For the calibration of RF Power Sensors - Calibration Factor excludes VRC measurements above 18 GHz.
3.  $f$  = frequency in hertz.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. L2448.



Jason Stine, Vice President

